

EC DECLARATION AND MANUFACTURER'S DECLARATION OF CONFORMITY

Manufacturer: Medela, Inc.
1101 Corporate Drive, McHenry, IL 60050 USA

EC Representative: Medela AG
Lättichstrasse 4b, CH-6341 Baar, Switzerland

Product Number/Description: 020.0000 / Calesca

Directives Applied: Low Voltage Directive 2006/95/EC
EMC Directive 2004/108/EC

Standards Applied: **ELECTRICAL SAFETY:**

EN 61010-1:2010- Safety Requirements for electrical equipment for measurement, control and laboratory use—Part 1: General Requirements.

EN 61010-2-010:2014 - Safety Requirements for electrical equipment for measurement, control and laboratory use—Part 2-010: Particular Requirements for laboratory equipment for the heating of material.


EMISSIONS:

EN 55011 :2009 + A1 :2010- Industrial, scientific and medical equipment - Radio-frequency disturbance characteristics - Limits and methods of measurement

EN 61326-1:2013- Electrical equipment for measurement, control and laboratory use - EMC requirements -- Part 1: General requirements

IMMUNITY:

EN 61000-4-2:2009, EN61000-4-3:2006 + A1:2006 + A2:2008, EN 61000-4-4:2004 + A1:2010, EN61000-4-5:2006, EN61000-4-6:2009, EN61000-4-8:2010, EN61000-4-11:2004



Robert Sokolowski
Director, Quality Assurance

Place of Signing: McHenry, IL 60050 USA

Date: June 16, 2016